



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/813,449	03/30/2004	Enrique J. Klein	16497.28.1	7612
57360 7590 08/14/2008 WORKMAN NYDEGGER 1000 EAGLE GATE TOWER, 60 EAST SOUTH TEMPLE SALT LAKE CITY, UT 84111				
EXAMINER				
DOWE, KATHERINE MARIE				
ART UNIT		PAPER NUMBER		
3734				
MAIL DATE		DELIVERY MODE		
08/14/2008		PAPER		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

### Office Action Summary

**Application No.**

10/813,449

**Applicant(s)**

KLEIN ET AL.

**Examiner**

KATHERINE M. DOWE

**Art Unit**

3734

**Period for Reply** -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 27 May 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1, 2, 5-14, 16-20, 23 and 24 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1, 2, 5-14, 16-20, 23, and 24 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO/55/00)  
Paper No(s)/Mail Date 10/23/2007 and 3/17/2008
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

**DETAILED ACTION**

1. The following is a complete response to the amendment filed 5/27/2008.
2. Claims 1, 2, 5-14, 16-20, 23, and 24 are currently pending.

***Claim Rejections - 35 USC § 112***

3. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

4. Claim 5 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. Claim 1 defines a port as being formed by advancing a needle through the tissue wall. A second port is not defined. Thus, it is unclear how the port in claim 5 is "on opposite sides of the aperture" (line 3). It appears two ports are being claimed for two distinct needles of the pair of needles because a single port can only be on a single side of the aperture.
5. Claim 6 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. Claim 1 defines a needle in line 3. Claim 6 recites "at least one needle having a portion of suture attached". It is unclear if the at least one needle in claim 6 is an additional needle or is the same needle positively recited in claim 1.

***Claim Rejections - 35 USC § 102***

6. The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.
7. Claims 1 and 2 are rejected under 35 U.S.C. 102(e) as being anticipated by Heidmueller (US 5,320,632). Heidmueller discloses a method for suturing tissue in the proximity of an aperture in a tissue wall comprising advancing a shaft (10) through the aperture (32) so that a pair of needles (25) is advanced from a distal side (Fig 2) of the tissue wall (33) to form a pair of ports with in the tissue wall on opposite sides of the aperture. At least a portion of suture (30) is passed from the distal side of the tissue wall proximally through the ports in the tissue wall (Figs 2-5; col 4, ln 14-18). A loop is formed with the remaining portion of suture to secure the suture (col 5, ln 3-4). The pair of needles may carry the suture portion therebetween as a continuous length (col 3, ln 65-67).

***Response to Arguments***

8. Applicant's arguments, see amendment, filed 5/27/2008, with respect to the rejection(s) of claim(s) 1, 2, 5-14, 16-20, 23, and 24 under Hathaway (US 5,304,184) have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made in view of Heidmueller (US 5,320,632).

***Conclusion***

9. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

10. Any inquiry concerning this communication or earlier communications from the examiner should be directed to **KATHERINE M. DOWE** whose telephone number is (571)272-3201. The examiner can normally be reached on M-F 8:30am - 5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jackie Ho can be reached on (571) 272-4696. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Kevin T. Truong/  
Primary Examiner, Art Unit 3734

Katherine Dowe  
August 10, 2008

/K. M. D./  
Examiner, Art Unit 3734